



CALIBRATION CERTIFICATE

CUSTOMER: Viper Rentals

DATE: 7-Dec-09

S.O. #: 13885

MAKE: 4.5" Base

S/N: VRS-1335

MODEL: CPW

P.O. #:

<u>Should Be Foot Pounds</u>	<u>Test Gauge PSI Reading</u>	<u>Actual Foot Pounds</u>
0	0	0
1000	250	1000
2000	500	2000
3000	750	3000
4000	1000	4000
5000	1250	5000
6000	1500	6000

Tech: C. Lindsay

Notes: New CPW

Test Gauge S/N: JW050830-7

We hereby certify, that on the above date, we repaired and re-calibrated
the above instrument and it is calibrated to National Standards.

506-23 A Avenue
Nisku AB T9E 8G2

GAUGE PLUS

Ph: 780-955-1797
Fx: 780-955-1796