



Calibration Certificate

CUSTOMER: Viper Rentals

DATE: 16-Feb-05

W.O. #:

MAKE: AIR

S/N:

MODEL: Torque Alert

P.O. #:

<u>Should Be Foot Pounds</u>	<u>Test Gauge PSI Reading</u>
0	0
1000	250
2000	500
3000	750
4000	1000
5000	1250
6000	1500

Tech: Mike Sine

Notes: 4.08 Cylinder

We hereby certify, that on the above data, we repaired and re-calibrated instrument and it is calibrated to National Standards.

Test Gauge S/N: JW20000926-5